

# Performance Testing of the System for Analysis of PADC Track Detectors for Neutron Dosimetry

Vendula Filová,<sup>1, a)</sup> Branislav Vrban,<sup>1, b)</sup> Štefan Čerba,<sup>1, c)</sup> Jakub Luley,<sup>1, d)</sup>  
Vladimír Nečas<sup>1, e)</sup> and Maria Dugdale<sup>2, f)</sup>

<sup>1</sup> *Slovak University of Technology in Bratislava, Faculty of Electrical Engineering and Information Technology, Institute of Nuclear and Physical Engineering, Ilkovičova 3, 81219 Bratislava, Slovakia.*

<sup>2</sup> *GM Scientific Ltd, Unit 12 Market Industrial Estate, North End Road, Yatton, BS49 4RF, United Kingdom.*

<sup>a)</sup> *Corresponding author: vendula.filova@stuba.sk*

<sup>b)</sup> *branislav.vrban@stuba.sk, <sup>c)</sup> stefan.cerba@stuba.sk, <sup>d)</sup> jakub.luley@stuba.sk,*

<sup>e)</sup> *vladimir.necas@stuba.sk, <sup>f)</sup> maria@gmscientific.uk*

**Abstract.** The new solid state track detector analysis system for neutron dosimetry and radon measurement TASLImage™ was installed at the Slovak University of Technology, Bratislava. This paper describes the test of its performance in the measurement of neutron doses. The linearity of the response of poly allyl diglycol carbonate (PADC) detectors was tested for two regions of obtained doses, 27- 378  $\mu$ Sv and 0.5 - 3 mSv. Results were compared to H\*(10) ambient dose equivalent measured by the NuDET-ENE neutron detector.